


148.00 mm

ZO-6065-6066

210.00 mm



AQUAPUMP

USER GUIDE

GETTING STARTED

For a tight seal, ensure your pubic hair is trimmed or shaved. Only use the minimal amount of pressure necessary for you to achieve an erection as over-pumping may result in injury, bruising, or damage to your blood vessels. Pump no more than 15 minutes for each 24 hour period. The pump may be used in & out of the bath or shower, as a waterpump or as a traditional air pump.

HOW TO USE IN THE SHOWER

Begin by pushing down on the pip and flicking it into the "closed" position. Proceed to fill the pump with water, then insert the penis and push the pump against the body to create a tight seal. Flick the pip to the "open" position and begin pumping until penis is erect. Push down on the release valve to reduce pressure and remove the pump.



Push and flick to "closed" position Push against body for tight seal Flick open pip and begin pumping Push down on valve to release pressure

HOW TO USE IN THE BATHTUB

Submerge the pump in water and let it fill, making sure that the hand pump remains above water. Flick the pip to the "closed" position, then insert the penis and proceed to push the pump against the body to create a tight seal. Flick the pip to the "open" position and begin pumping until penis is erect. Push down on the release valve to reduce pressure and remove the pump.

HOW TO USE WITHOUT THE HAND-PUMP

The ZOLO Aquapump may be used without the hand pump. Detach the hand pump's tube from the release valve, then cover the top of the valve with a finger. Submerge the pump in water and let it fill, then insert the penis and proceed to push the pump against the body to create a tight seal. Remove the finger from the release valve and begin pumping against your body. Pump until penis is erect then push down the release valve to reduce pressure, before finally removing the pump.



WARNING!

- Cease usage of the vacuum pump if pain occurs.
- Avoid the usage of the vacuum pump while under the influence of alcohol or drugs, as this may impair the user's judgement and increase the risk of penis injury.
- Do not use if the user has sickle cell disease, a history of prolonged erections, or is taking large quantities of aspirin or other blood thinners. These conditions may increase the risk of bruising and hematomas.
- Misuse of the vacuum pump may cause penis injury.
- If any discomfort occurs during use, discontinue use of pump immediately.

CAUTION

- Avoid over-vigorous pumping.
- This product is intended for the use by a single user.
- Cannot be used for longer than 15 minutes in a 24 hour period.
- Do not fall asleep while using the vacuum pump, or while under the influence of alcohol or drugs. The user needs to be in control of the product at all times.
- Consult a doctor before using the vacuum pump if the user experiences any discomfort in the genital area.
- The user must be in full control of his hands to safely use the device with the strength to push the pressure release button.

DEVICE PURPOSE

- The device is an external penile rigidity device using vacuum technology.
- The device is also intended to maintain optimum penis health and sexual stamina, through regular exercise of the penis by erection using a safe level of vacuum pressure.

The following precautionary measures are advised when using the device.

As the device may cause sexual arousal with increased heart and respiratory rates, users should consult their doctors before using the device if they suffer any of the following: heart or lung disease, high blood pressure, stroke, epilepsy or convulsant disorder, uncontrollable diabetes or genital-urinary disorder (kidney, bladder, testicular or penile problems etc). Consult a doctor before use if user consumes anticoagulants or have sickle cell anaemia, bleeding disorders or leukaemia.

PRECAUTIONS FOR CONSTRICTION RINGS AND VACUUM PUMPS

DELAYED DIAGNOSIS OF OTHER CONDITIONS

If the user has symptoms of erectile dysfunction (i.e. inability to achieve an erection that is sufficient for sexual intercourse) consult a doctor prior to using this device to avoid any potentially harmful delay in diagnosing any of the most common causes of this condition.

USE WITH IMPAIRED PAIN PERCEPTION

Do not use the device if the user has decreased sensation of pain in the penis region. This warning should inform the user if he cannot feel pain, he may not know if the device is causing him harm or injury.

USE WITH DECREASED HAND STRENGTH

Do not use the device if the user has decreased hand strength as this may make the removal of the device difficult.

DISCLAIMER

User acknowledges and agrees that although this product contains pumps in general are safe for most men, there are some potential risks, which could be serious. As such, user acknowledges and agrees to discuss the use of this product with their doctor before use and should tell your doctor about all of your health conditions and any medications you take, including herbal supplements. This will help prevent potential problems. According to the Food and Drug Administration (FDA) Trusted Source, using a penis pump or other external rigidity device could aggravate certain penile conditions and must be discussed with your doctor. Use only as directed. User releases manufacturer and/or seller for any liability related to use of this product and cannot make claims or sue for personal injury.

SOLD AS A NOVELTY ONLY
WWW.ZOLOCUP.COM

UK Responsible Person:

ArangaRanga UK LTD, 20-22 Wenlock Road,
London, N17 GU, UK.
Phone: +44 2037692960
E-Mail: UKresponsibleperson@arangaranga.com

EU Responsible Person:

ArangaRanga SL, Paseo de la Castellana 194,
Madrid, 28046, Spain.
Phone: +34910054258
E-Mail: EUresponsibleperson@arangaranga.com

Distributed by XGEN, LLC
Horsham, PA 19044

MADE IN CHINA



Problem	Cause	Solution	Notes
No erection	No vacuum due to air leak from poor seal	- Use more lubricant - Apply firm pressure to body - Trim pubic hair - Check device for air leaks	Problems are most often associated with inexperience from using the device
Partial erection	- Incorrect insert or seal ring size - Inexperience in operating the device	- More practice from using device - Use larger cylinder insert	Problems are most often associated with inexperience in operating the device
Rapid loss of erection (no more than 5 minutes)	- Vacuum loss due to air leak from faulty equipment - Poor seal - Inadequate lubrication - Excessive pubic hair	- Use more lubricant - Apply firm pressure to body - Trim pubic hair - Check device for air leaks - Use smaller cylinder insert - Use smaller or double tension rings - Remove tension band carefully	- Problems are most often associated with inexperience from using the device - Consult a physician if problem occurs
Pulling of the skin of the scrotum	- Insufficient tension from bands or rings	- Remove lubricant from scrotum - Angle the device downwards when in use	- Problems are most often associated with inexperience from using the device
Experiencing any discomfort or pain while device is in use	- Pumping too rapidly - Pulling of scrotal tissue	- Slower pumping - Continue to use device	- Problems are most often associated with inexperience from using the device - Consult a physician if problem occurs
Having redness, irritation or bruising	- Pumping too rapidly - Over pumping	- Slower pumping - Fewer pumps	- Problems are most often associated with inexperience from using the device - Consult a physician if problem occurs
Feeling a cold penis (to you and your partner)	- Constriction of blood flow to the penis - Lubrication	- Wash off lubricant or warm lubricant before use	-
Penis pivots on its base	- Constriction of blood flow to the penis	- If manageable, apply the device after you have a partial erection	-

Front

[Back](#)